

## GOVERNMENT OF WEST BENGAL OFFICE OF THE PRINCIPAL DEBEN MAHATA GOVERNMENT MEDICAL COLLEGE & HOSPITAL

Vill.:Hatuara, P.O.:Vivekananda Nagar, P.S.:PuruliaMuffasil,PIN: 723 147 dmgmch.edu.in: prinpuruliagmch@gmail.com

Memo No: DMGMCH /PRL/

Dated: July

, 2023

## **QUOTATION NOTICE FOR EMBALMING MACHINE**

Sealed quotations are hereby invited from reputed firms/vendors/direct importers/ authorized distributor firms for supply & installation of **Embalming Machine & Other Accessories** for the Department of Anatomy of Deben Mahata Government Medical College & Hospital.

Details of the items are as specified in Annexure - I

The last date of submission of Quotation is **27.07.2023** and same will be opened on the same day at 3:30 pm in presence of interested bidders, if any. The bidders will submit the quotations in their own official pad along with the following essential documents:

- 1. Valid Trade License / Enlistment
- 2. Pan Card of Bidder/ Bidders Company
- 3. Manufacturing License (In Case of Manufacture)
- 4. GST Registration Certificate
- 5. Last Quarters' GST Return
- 6. IT Returns of last year
- 7. Export Import license with IEC code (For Importer)
- 8. The certificate in the category of BIS/IS/FDA/US FDA of European CE with registration number of the product or Products (if applicable)
- 9. Brochure of the product or Products
- 10. Previous credentials of supplying similar products in reputed Medical Colleges (preferably Government Medical College)

The Quotation is to be submitted in the drop box which will be kept at the Office/Establishment Section of the undersigned on all working days in working hours from 11 am to 3 pm or may be sent by post within the due date.

Physical verification of samples and/ or physical demonstration may be done at the discretion of the hospital authority by a team of experts to adjudge the applicability or suitability of the product for the functional requirements of the Department and the technical evaluation will be done accordingly in close observation of the product or products. The undersigned reserves the right to accept or cancel the whole procedure at any time without showing any reason.

Principal

Deben Mahata Government Medical College & Hospital

Purulia

## Memo No: DMGMCH /PRL/ 734 /1(12)

Copy forwarded for information & necessary action to:

- 1. Sabhadhipati, Purulia Zilla Parishad
- 2. The Director of Medical Education, Department of Health & Family Welfare, Government of West Bengal, Swasthya Bhawan, Salt Lake City, Kolkata - 91
- 3. The District Magistrate, Purulia
- 4. The MSVP, DMGMC&H, Purulia
- 5. The Chief Medical Officer of Health, Purulia
- 6. DICO, Purulia
- 7. The IT Cell, Swasthya Bhawan, Kol-91 for uploading this Quotation notice in the official website of Dept. of Health & Family Welfare
- 8. District Information Officer, NIC, Purulia with request to publish the Quotation through official website of Purulia District
- 9. The Accounts Officer, DMGMCH
- 10. Notice Board of DMGMCH& DMSH
- 11. DMGMCH Website

12. Office copy

Principal College & Hospital

Purulia (Indica) Purulia

Purulia (Indica) Purulia

Perentalia Containa Purulia

Perentalia Containa Purulia

## Annexure – I Specification

Sl. No.	Configuration Details	Quantity
1	A completely silent & vibration free machine used for injecting Formaldehyde solution in cadavers at much higher speed than gravity process	
2	The machine is to be mounted inside a covered cabinet fitted with castor wheels for ease in movement	
3	The unit should consist of a suitable air compressor fitted with an appropriate motor (0.5 hp to 1 hp) which should be connected with a stainless tank of 10 litres capacity meant for storing and injecting the solution	01 No.
4	The tank should be fitted with a safety valve, pressure gauge and rubber tubing having provision for injection	
5	The set should be provided with electric cord, plug and allied to work on 220V/50 Hz A.C. supply	<u> </u>
6	Straight cannula to be provided with the machine	Big size – 02 Pcs Small size-02 Pcs
7	Appropriate warranty of minimum 3 years	