

GOVERNMENT OF WEST BENGAL OFFICE OF THE PRINCIPAL

PURULIA GOVERNMENT MEDICAL COLLEGE & HOSPITAL

Vill.:Hatuara, P.O.:Vivekananda Nagar, P.S.:Purulia Muffasil,PIN: 723 147 pgmch.edu.in: prinpuruliagmch@gmail.com

Memo.No: PGMCH / PRL/ 179

Dated: February 11,2020

QUOTATION NOTICE

Sealed quotations are hereby invited from reputed firms/ direct importers/ authorized distributors firm for supply of **Embalming Machine (Cadaver Injector)** for use in the **Department of Anatomy.**

Detailed specifications are as specified in Annexure I.

The last date of submission of Tender is **19.02.2020** and same will be opened on the same day at 4.00 pm in presence of interested bidders, if any. The bidders will submit the quotations in their own official pad along with the following essential documents:

- 1. Valid Trade License / Enlistment
- 2. Pan Card of Bidder/ Bidders Company
- 3. Manufacturing License (In Case of Manufacture)
- 4. GST Registration Certificate
- 5. Last Quarters' GST Return
- 6. IT Returns of last year
- 7. Export Import license with IEC code (For Importer)
- 8. The certificate in the category of BIS/IS/FDA/US FDA of European CE with registration number of the product or Products (if applicable)
- 9. Brochure of the product or Products
- 10. Previous credentials of supplying similar products in reputed Medical Colleges (preferably Government Medical College)

The Quotation is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days in working hours from 11 am to 3 pm or may be sent by post within the due date.

Physical verification of samples and/ or physical demonstration may be done at the discretion of the hospital authority by a team of experts to adjudge the applicability or suitability of the product for the functional requirements of the Department and the technical evaluation will be done accordingly in close observation of the product or products. The undersigned reserves the right to accept or cancel the whole procedure at any time without showing any reason.

Principal Purulia Government Medical College & Hospital

Purulia

Memo.No: PGMCH /PRL/

Copy forwarded for information & necessary action to:

- 1. Sabhadipati, Purulia Zila Parishad
- The Director of Medical Education, Department of Health & Family Welfare, Government of West Bengal, Swasthya Bhawan, Salt Lake City, Kolkata – 91
- 3. The District Magistrate, Purulia
- 4. The MSVP, PGMC&H, Purulia
- 5. The Chief Medical Officer of Health, Purulia
- 6. DICO, Purulia
- 7. The IT Cell, Swasthya Bhawan, Kol-91 for uploading this Quotation notice in the official website of Dept. of Health & Family Welfare
- 8. District Information Officer, NIC, Purulia with request to publish the Quotation through official website of Purulia District
- 9. The Accounts Officer, PGMCH
- 10. Notice Board of PGMCH & DMSH
- 11. PGMCH Website
- 12. Office copy

Principal Purulia Government Medical College & Hospital Purulia

<u> Annexure – I</u>

Specifications of Embalming Machine (Cadaver Injector)

- 1. A completely silent & vibration free machine used for injecting Formaldehyde solution in Cadaverous at much higher speed than normal gravity process.
- 2. The machine is to be mounted inside a covered cabinet fitted with castor wheels for ease in movement.
- 3. The Unit should consist of a suitable air compressor fitted with an appropriate motor (1/2 hp to 1 hp) which should be connected with a stainless tank of 10 litres capacity meant for storing and injecting the solution.
- 4. The tank should be fitted with a safety valve, pressure gauge and rubber tubing having provision for injection.
- 5. The Set should be provided with electric cord, plug and allied to work on 220V/50 Hz A.C. supply.
- 6. Four Straight Cannula Two Big Size & Two Small Size are to be provided with the machine.
- 7. Appropriate Warranty of minimum 3 Years

ectimican menure west ATUARA, PURULA, WEST